

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY111122\
 Data File : VY011414.D
 Acq On : 11 Nov 2022 13:31
 Operator : KP/MD
 Sample : N5530-03
 Misc : 5.05g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SB-12

Quant Time: Nov 11 13:47:15 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y111022S.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 10 17:47:26 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	335391	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	641096	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	600266	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	285297	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	214619	76.364	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	152.720%
35) Dibromofluoromethane	7.716	113	213604	59.453	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	118.900%
50) Toluene-d8	10.179	98	673922	49.099	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	98.200%
62) 4-Bromofluorobenzene	12.477	95	258952	54.090	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	108.180%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY111122\
Data File : VY011414.D
Acq On : 11 Nov 2022 13:31
Operator : KP/MD
Sample : N5530-03
Misc : 5.05g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SB-12

Quant Time: Nov 11 13:47:15 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y111022S.M
Quant Title : SW846 8260
QLast Update : Thu Nov 10 17:47:26 2022
Response via : Initial Calibration

