

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY111122\
 Data File : VY011428.D
 Acq On : 11 Nov 2022 18:49
 Operator : KP/MD
 Sample : N5580-01
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SU-02-111022

Quant Time: Nov 14 05:54:29 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y111022S.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 10 17:47:26 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	47542	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	75742	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	62327	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	22781	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	32348	81.197	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	162.400%
35) Dibromofluoromethane	7.716	113	31795	74.904	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	149.800%#
50) Toluene-d8	10.179	98	92657	57.138	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	114.280%
62) 4-Bromofluorobenzene	12.477	95	26281	46.465	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	92.920%
Target Compounds						Qvalue
86) p-Isopropyltoluene	13.367	119	4054	2.861	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY111122\
Data File : VY011428.D
Acq On : 11 Nov 2022 18:49
Operator : KP/MD
Sample : N5580-01
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SU-02-111022

Quant Time: Nov 14 05:54:29 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y111022S.M
Quant Title : SW846 8260
QLast Update : Thu Nov 10 17:47:26 2022
Response via : Initial Calibration

