

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY111122\
 Data File : VY011430.D
 Acq On : 11 Nov 2022 19:34
 Operator : KP/MD
 Sample : N5583-01
 Misc : 4.99g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 NB-08-111022

Quant Time: Nov 14 05:54:47 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y111022S.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 10 17:47:26 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	343252	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	576969	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	509707	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	229311	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	174227	60.572	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	121.140%
35) Dibromofluoromethane	7.716	113	187635	58.029	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	116.060%
50) Toluene-d8	10.179	98	700132	56.678	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	113.360%
62) 4-Bromofluorobenzene	12.477	95	221941	51.512	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	103.020%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY111122\
Data File : VY011430.D
Acq On : 11 Nov 2022 19:34
Operator : KP/MD
Sample : N5583-01
Misc : 4.99g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 28 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
NB-08-111022

Quant Time: Nov 14 05:54:47 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y111022S.M
Quant Title : SW846 8260
QLast Update : Thu Nov 10 17:47:26 2022
Response via : Initial Calibration

