

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY111124\
 Data File : VY020248.D
 Acq On : 11 Nov 2024 15:12
 Operator : SY/MD
 Sample : P4768-06
 Misc : 2.32g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 B-6-2

Quant Time: Nov 11 23:16:49 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y103024S.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 31 15:23:13 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.713	168	190237	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.616	114	399773	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.414	117	365761	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.346	152	128461	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.061	65	135459	57.053	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	114.100%
35) Dibromofluoromethane	7.634	113	132702	52.209	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	104.420%
50) Toluene-d8	10.109	98	495141	48.940	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	97.880%
62) 4-Bromofluorobenzene	12.408	95	151247	46.024	ug/l	0.00
Spiked Amount	50.000	Range	29 - 146	Recovery	=	92.040%
Target Compounds						
68) m/p-Xylenes	11.621	106	10308	1.934	ug/l	Qvalue 93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY111124\
Data File : VY020248.D
Acq On : 11 Nov 2024 15:12
Operator : SY/MD
Sample : P4768-06
Misc : 2.32g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
B-6-2

Quant Time: Nov 11 23:16:49 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y103024S.M
Quant Title : SW846 8260
QLast Update : Thu Oct 31 15:23:13 2024
Response via : Initial Calibration

