

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY111124\
 Data File : VY020251.D
 Acq On : 11 Nov 2024 16:22
 Operator : SY/MD
 Sample : P4788-03
 Misc : 4.12g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 BP-G3-VOC

Quant Time: Nov 11 23:18:00 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y103024S.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 31 15:23:13 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.713	168	176161	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.616	114	382066	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.414	117	352320	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.346	152	126271	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.061	65	128332	58.370	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 116.740%	
35) Dibromofluoromethane	7.634	113	124943	51.435	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 102.860%	
50) Toluene-d8	10.109	98	475150	49.140	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	= 98.280%	
62) 4-Bromofluorobenzene	12.408	95	147420	46.939	ug/l	0.00
Spiked Amount	50.000	Range	29 - 146	Recovery	= 93.880%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY111124\
Data File : VY020251.D
Acq On : 11 Nov 2024 16:22
Operator : SY/MD
Sample : P4788-03
Misc : 4.12g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
BP-G3-VOC

Quant Time: Nov 11 23:18:00 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y103024S.M
Quant Title : SW846 8260
QLast Update : Thu Oct 31 15:23:13 2024
Response via : Initial Calibration

