

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY111124\
 Data File : VY020252.D
 Acq On : 11 Nov 2024 16:46
 Operator : SY/MD
 Sample : P4789-03
 Misc : 5.11g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 BP-F21-VOC

Quant Time: Nov 11 23:18:20 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y103024S.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 31 15:23:13 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.713	168	119529	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.616	114	234648	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.414	117	209503	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.347	152	69599	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.061	65	81689	54.759	ug/l	0.00
Spiked Amount	50.000	Range 50 - 163	Recovery	= 109.520%		
35) Dibromofluoromethane	7.640	113	77122	51.695	ug/l	0.00
Spiked Amount	50.000	Range 54 - 147	Recovery	= 103.380%		
50) Toluene-d8	10.109	98	287553	48.423	ug/l	0.00
Spiked Amount	50.000	Range 58 - 134	Recovery	= 96.840%		
62) 4-Bromofluorobenzene	12.408	95	82536	42.790	ug/l	0.00
Spiked Amount	50.000	Range 29 - 146	Recovery	= 85.580%		
Target Compounds						
16) Acetone	3.879	43	7346	21.175	ug/l	Qvalue 93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY111124\
Data File : VY020252.D
Acq On : 11 Nov 2024 16:46
Operator : SY/MD
Sample : P4789-03
Misc : 5.11g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
BP-F21-VOC

Quant Time: Nov 11 23:18:20 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y103024S.M
Quant Title : SW846 8260
QLast Update : Thu Oct 31 15:23:13 2024
Response via : Initial Calibration

