

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY111124\
 Data File : VY020239.D
 Acq On : 11 Nov 2024 11:11
 Operator : SY/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 LabSampleId :
 VSTDCCC050

Quant Time: Nov 11 23:11:07 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y103024S.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 31 15:23:13 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	101	0.00
2 T	Dichlorodifluoromethane	0.463	0.427	7.8	114	0.00
3 P	Chloromethane	0.767	0.726	5.3	108	0.00
4 C	Vinyl Chloride	0.823	0.857	-4.1#	115	0.00
5 T	Bromomethane	0.546	0.543	0.5	115	0.00
6 T	Chloroethane	0.531	0.560	-5.5	117	0.00
7 T	Trichlorofluoromethane	1.005	1.112	-10.6	122	0.00
8 T	Diethyl Ether	0.280	0.310	-10.7	125	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.564	0.594	-5.3	118	0.00
10 T	Methyl Iodide	0.544	0.607	-11.6	119	0.00
11 T	Tert butyl alcohol	0.045	0.040	11.1	124	0.00
12 CM	1,1-Dichloroethene	0.540	0.581	-7.6#	120	0.00
13 T	Acrolein	0.056	0.034	39.3#	69	0.00
14 T	Allyl chloride	0.973	1.044	-7.3	116	0.00
15 T	Acrylonitrile	0.128	0.134	-4.7	116	0.00
16 T	Acetone	0.145	0.173	-19.3	125	0.00
17 T	Carbon Disulfide	1.796	1.791	0.3	108	0.00
18 T	Methyl Acetate	0.340	0.362	-6.5	116	0.00
19 T	Methyl tert-butyl Ether	1.357	1.524	-12.3	125	0.00
20 T	Methylene Chloride	0.652	0.616	5.5	113	0.00
21 T	trans-1,2-Dichloroethene	0.609	0.647	-6.2	116	0.00
22 T	Diisopropyl ether	2.020	2.214	-9.6	119	0.00
23 T	Vinyl Acetate	1.137	1.220	-7.3	117	0.00
24 P	1,1-Dichloroethane	1.170	1.268	-8.4	120	0.00
25 T	2-Butanone	0.197	0.215	-9.1	122	0.00
26 T	2,2-Dichloropropane	0.973	1.082	-11.2	125	0.00
27 T	cis-1,2-Dichloroethene	0.700	0.774	-10.6	121	0.00
28 T	Bromochloromethane	0.548	0.576	-5.1	116	0.00
29 T	Tetrahydrofuran	0.114	0.120	-5.3	117	0.00
30 C	Chloroform	1.161	1.287	-10.9#	122	0.00
31 T	Cyclohexane	1.132	1.098	3.0	113	0.00
32 T	1,1,1-Trichloroethane	1.013	1.141	-12.6	122	0.00
33 S	1,2-Dichloroethane-d4	0.624	0.635	-1.8	110	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	101	0.00
35 S	Dibromofluoromethane	0.318	0.337	-6.0	111	0.00
36 T	1,1-Dichloropropene	0.491	0.534	-8.8	120	0.00
37 T	Ethyl Acetate	0.234	0.246	-5.1	118	0.00
38 T	Carbon Tetrachloride	0.494	0.562	-13.8	123	0.00
39 T	Methylcyclohexane	0.623	0.655	-5.1	114	0.00
40 TM	Benzene	1.433	1.567	-9.4	119	0.00
41 T	Methacrylonitrile	0.117	0.140	-19.7	118	0.00
42 TM	1,2-Dichloroethane	0.397	0.429	-8.1	119	0.00
43 T	Isopropyl Acetate	0.450	0.483	-7.3	118	0.00
44 TM	Trichloroethene	0.332	0.369	-11.1	121	0.00
45 C	1,2-Dichloropropane	0.343	0.375	-9.3#	119	0.00
46 T	Dibromomethane	0.186	0.201	-8.1	118	0.00
47 T	Bromodichloromethane	0.487	0.554	-13.8	124	0.00
48 T	Methyl methacrylate	0.209	0.227	-8.6	116	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY111124\
 Data File : VY020239.D
 Acq On : 11 Nov 2024 11:11
 Operator : SY/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 LabSampleId :
 VSTDCCC050

Quant Time: Nov 11 23:11:07 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y103024S.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 31 15:23:13 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.002	0.002	0.0	121	0.00
50 S	Toluene-d8	1.265	1.329	-5.1	110	0.00
51 T	4-Methyl-2-Pentanone	0.232	0.255	-9.9	117	0.00
52 CM	Toluene	0.884	0.983	-11.2#	119	0.00
53 T	t-1,3-Dichloropropene	0.430	0.488	-13.5	123	0.00
54 T	cis-1,3-Dichloropropene	0.513	0.576	-12.3	120	0.00
55 T	1,1,2-Trichloroethane	0.231	0.260	-12.6	122	0.00
56 T	Ethyl methacrylate	0.331	0.379	-14.5	122	0.00
57 T	1,3-Dichloropropane	0.423	0.466	-10.2	120	0.00
58 T	2-Chloroethyl Vinyl ether	0.163	0.176	-8.0	111	0.00
59 T	2-Hexanone	0.166	0.188	-13.3	121	0.00
60 T	Dibromochloromethane	0.295	0.342	-15.9	127	0.00
61 T	1,2-Dibromoethane	0.215	0.234	-8.8	121	0.00
62 S	4-Bromofluorobenzene	0.411	0.447	-8.8	113	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	101	0.00
64 T	Tetrachloroethene	0.330	0.382	-15.8	125	0.00
65 PM	Chlorobenzene	1.065	1.193	-12.0	121	0.00
66 T	1,1,1,2-Tetrachloroethane	0.340	0.398	-17.1	126	0.00
67 C	Ethyl Benzene	1.992	2.238	-12.3#	120	0.00
68 T	m/p-Xylenes	0.729	0.824	-13.0	120	0.00
69 T	o-Xylene	0.690	0.780	-13.0	121	0.00
70 T	Styrene	1.141	1.333	-16.8	123	0.00
71 P	Bromoform	0.175	0.203	-16.0	124	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	100	0.00
73 T	Isopropylbenzene	3.994	4.485	-12.3	122	0.00
74 T	N-amyl acetate	0.998	1.119	-12.1	122	0.00
75 P	1,1,2,2-Tetrachloroethane	0.699	0.743	-6.3	119	0.00
76 T	1,2,3-Trichloropropane	0.469	0.536	-14.3	125	0.00
77 T	Bromobenzene	0.808	0.888	-9.9	119	0.00
78 T	n-propylbenzene	4.971	5.491	-10.5	120	0.00
79 T	2-Chlorotoluene	2.759	3.032	-9.9	120	0.00
80 T	1,3,5-Trimethylbenzene	3.207	3.563	-11.1	120	0.00
81 T	trans-1,4-Dichloro-2-butene	0.218	0.241	-10.6	122	0.00
82 T	4-Chlorotoluene	2.813	3.149	-11.9	121	0.00
83 T	tert-Butylbenzene	2.804	3.233	-15.3	121	0.00
84 T	1,2,4-Trimethylbenzene	3.176	3.553	-11.9	121	0.00
85 T	sec-Butylbenzene	4.305	4.835	-12.3	121	0.00
86 T	p-Isopropyltoluene	3.454	3.869	-12.0	119	0.00
87 T	1,3-Dichlorobenzene	1.641	1.812	-10.4	120	0.00
88 T	1,4-Dichlorobenzene	1.634	1.772	-8.4	119	0.00
89 T	n-Butylbenzene	3.463	3.853	-11.3	118	0.00
90 T	Hexachloroethane	0.683	0.777	-13.8	123	0.00
91 T	1,2-Dichlorobenzene	1.426	1.583	-11.0	121	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.106	0.116	-9.4	126	0.00
93 T	1,2,4-Trichlorobenzene	0.766	0.851	-11.1	123	0.00
94 T	Hexachlorobutadiene	0.419	0.457	-9.1	117	0.00
95 T	Naphthalene	1.432	1.661	-16.0	125	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY111124\
Data File : VY020239.D
Acq On : 11 Nov 2024 11:11
Operator : SY/MD
Sample : VSTDCCC050
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_Y
LabSampleId :
VSTDCCC050

Quant Time: Nov 11 23:11:07 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y103024S.M
Quant Title : SW846 8260
QLast Update : Thu Oct 31 15:23:13 2024
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.650	0.711	-9.4	123	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6