



Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY111124\
Data File : VY020238.D
Acq On : 11 Nov 2024 09:30
Operator : SY/MD
Sample : BFB
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y103024S.M
Title : SW846 8260
Last Update : Thu Oct 31 15:23:13 2024



AutoFind: Scans 1752, 1753, 1754; Background Corrected with Scan 1744

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	24.2	39845	PASS
75	95	30	60	57.9	95432	PASS
95	95	100	100	100.0	164693	PASS
96	95	5	9	6.5	10769	PASS
173	174	0.00	2	1.4	1675	PASS
174	95	50	100	73.1	120309	PASS
175	174	5	9	7.4	8916	PASS
176	174	95	101	97.5	117272	PASS
177	176	5	9	6.0	7053	PASS