

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY111224\
 Data File : VY020266.D
 Acq On : 12 Nov 2024 14:38
 Operator : SY/MD
 Sample : P4779-01RE
 Misc : 4.99g/5.0mL/MSVOA_Y/SOIL/B
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SOIL-01RE

Quant Time: Nov 13 00:35:34 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y103024S.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 31 15:23:13 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.719	168	7140	50.000	ug/l	0.01
34) 1,4-Difluorobenzene	8.622	114	14385	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.426	117	10452	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.359	152	2638	50.000	ug/l	0.01
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.067	65	7590	85.174	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	170.340%#
35) Dibromofluoromethane	7.628	113	4878	53.335	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	106.680%
50) Toluene-d8	10.109	98	15982	43.900	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	87.800%
62) 4-Bromofluorobenzene	12.407	95	2709	22.909	ug/l	0.00
Spiked Amount	50.000	Range	29 - 146	Recovery	=	45.820%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY111224\
Data File : VY020266.D
Acq On : 12 Nov 2024 14:38
Operator : SY/MD
Sample : P4779-01RE
Misc : 4.99g/5.0mL/MSVOA_Y/SOIL/B
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SOIL-01RE

Quant Time: Nov 13 00:35:34 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y103024S.M
Quant Title : SW846 8260
QLast Update : Thu Oct 31 15:23:13 2024
Response via : Initial Calibration

