

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY111225\
 Data File : VY023764.D
 Acq On : 12 Nov 2025 15:21
 Operator : SY/MD
 Sample : Q3609-05
 Misc : 5.66g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 AU-713-VOC-02

Quant Time: Nov 14 03:39:44 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y110425S.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 05 04:41:50 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.707	168	734904	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.616	114	1269960	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.414	117	1079147	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.346	152	429003	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.061	65	356298	56.837	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	113.680%
35) Dibromofluoromethane	7.634	113	402421	50.706	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	101.420%
50) Toluene-d8	10.109	98	1420918	48.063	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	96.120%
62) 4-Bromofluorobenzene	12.408	95	415391	43.448	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	86.900%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY111225\
Data File : VY023764.D
Acq On : 12 Nov 2025 15:21
Operator : SY/MD
Sample : Q3609-05
Misc : 5.66g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
AU-713-VOC-02

Quant Time: Nov 14 03:39:44 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y110425S.M
Quant Title : SW846 8260
QLast Update : Wed Nov 05 04:41:50 2025
Response via : Initial Calibration

