

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY111225\
 Data File : VY023770.D
 Acq On : 12 Nov 2025 17:42
 Operator : SY/MD
 Sample : Q3614-03
 Misc : 5.42g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 COMP-3

Quant Time: Nov 14 03:41:50 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y110425S.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 05 04:41:50 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.707	168	759606	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.616	114	1324033	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.414	117	1144655	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.347	152	465011	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.061	65	385139	59.439	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	118.880%
35) Dibromofluoromethane	7.634	113	421681	50.963	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	101.920%
50) Toluene-d8	10.103	98	1505208	48.835	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	97.660%
62) 4-Bromofluorobenzene	12.402	95	449674	45.113	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	90.220%
Target Compounds						
16) Acetone	3.866	43	13136	8.156	ug/l	Qvalue 95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY111225\
Data File : VY023770.D
Acq On : 12 Nov 2025 17:42
Operator : SY/MD
Sample : Q3614-03
Misc : 5.42g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
COMP-3

Quant Time: Nov 14 03:41:50 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y110425S.M
Quant Title : SW846 8260
QLast Update : Wed Nov 05 04:41:50 2025
Response via : Initial Calibration

