

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY112024\
 Data File : VY020352.D
 Acq On : 20 Nov 2024 12:20
 Operator : SY/MD
 Sample : P4910-03
 Misc : 5.74g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 MH-COTTAGE-VOC

Quant Time: Nov 21 00:45:46 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y111924S.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 20 04:38:24 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.707	168	110405	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.616	114	215388	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.414	117	201860	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.346	152	80899	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.067	65	78000	61.494	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	122.980%
35) Dibromofluoromethane	7.634	113	69352	50.214	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	100.420%
50) Toluene-d8	10.109	98	261670	48.096	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	96.200%
62) 4-Bromofluorobenzene	12.402	95	91017	52.041	ug/l	0.00
Spiked Amount	50.000	Range	29 - 146	Recovery	=	104.080%

Target Compounds Qvalue

 (#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY112024\
Data File : VY020352.D
Acq On : 20 Nov 2024 12:20
Operator : SY/MD
Sample : P4910-03
Misc : 5.74g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
MH-COTTAGE-VOC

Quant Time: Nov 21 00:45:46 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y111924S.M
Quant Title : SW846 8260
QLast Update : Wed Nov 20 04:38:24 2024
Response via : Initial Calibration

