

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY112024\
 Data File : VY020365.D
 Acq On : 20 Nov 2024 17:24
 Operator : SY/MD
 Sample : P4849-01
 Misc : 5.62g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 RR-1

Quant Time: Nov 21 00:50:54 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y111924S.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 20 04:38:24 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.713	168	180085	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.616	114	325095	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.414	117	294030	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.346	152	110294	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.061	65	102609	49.594	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	99.180%
35) Dibromofluoromethane	7.640	113	97821	46.926	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	93.860%
50) Toluene-d8	10.109	98	378021	46.034	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	92.060%
62) 4-Bromofluorobenzene	12.408	95	124509	47.166	ug/l	0.00
Spiked Amount	50.000	Range	29 - 146	Recovery	=	94.340%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY112024\
Data File : VY020365.D
Acq On : 20 Nov 2024 17:24
Operator : SY/MD
Sample : P4849-01
Misc : 5.62g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
RR-1

Quant Time: Nov 21 00:50:54 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y111924S.M
Quant Title : SW846 8260
QLast Update : Wed Nov 20 04:38:24 2024
Response via : Initial Calibration

