

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY112024\
 Data File : VY020369.D
 Acq On : 20 Nov 2024 18:58
 Operator : SY/MD
 Sample : P4860-06
 Misc : 7.41g/5.0mL/MSVOA_Y/SOIL/B
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 PH2-BOT-009

Quant Time: Nov 21 00:52:29 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y111924S.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 20 04:38:24 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.713	168	142975	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.616	114	273649	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.414	117	253521	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.346	152	99327	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.061	65	84547	51.471	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	102.940%
35) Dibromofluoromethane	7.634	113	84465	48.136	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	96.280%
50) Toluene-d8	10.103	98	332194	48.059	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	96.120%
62) 4-Bromofluorobenzene	12.402	95	111020	49.963	ug/l	0.00
Spiked Amount	50.000	Range	29 - 146	Recovery	=	99.920%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY112024\
Data File : VY020369.D
Acq On : 20 Nov 2024 18:58
Operator : SY/MD
Sample : P4860-06
Misc : 7.41g/5.0mL/MSVOA_Y/SOIL/B
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
PH2-BOT-009

Quant Time: Nov 21 00:52:29 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y111924S.M
Quant Title : SW846 8260
QLast Update : Wed Nov 20 04:38:24 2024
Response via : Initial Calibration

