

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY112024\
 Data File : VY020370.D
 Acq On : 20 Nov 2024 19:22
 Operator : SY/MD
 Sample : P4768-04
 Misc : 4.50g/5.0mL/MSVOA_Y/SOIL/B
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 B-4

Quant Time: Nov 21 05:05:57 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y111924S.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 20 04:38:24 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.707	168	142436	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.615	114	269453	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.414	117	254267	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.346	152	98434	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.061	65	88311	53.966	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	107.940%
35) Dibromofluoromethane	7.634	113	86072	49.816	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	99.640%
50) Toluene-d8	10.109	98	332979	48.923	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	97.840%
62) 4-Bromofluorobenzene	12.407	95	110952	50.710	ug/l	0.00
Spiked Amount	50.000	Range	29 - 146	Recovery	=	101.420%

Target Compounds Qvalue

 (#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY112024\
Data File : VY020370.D
Acq On : 20 Nov 2024 19:22
Operator : SY/MD
Sample : P4768-04
Misc : 4.50g/5.0mL/MSVOA_Y/SOIL/B
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
B-4

Quant Time: Nov 21 05:05:57 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y111924S.M
Quant Title : SW846 8260
QLast Update : Wed Nov 20 04:38:24 2024
Response via : Initial Calibration

