

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY112223\
 Data File : VY016450.D
 Acq On : 22 Nov 2023 09:37
 Operator : SY/MD
 Sample : VY1122SBL01
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VY1122SBL01

Quant Time: Nov 23 04:02:17 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y110823S.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 09 00:45:52 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	100067	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	162560	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.496	117	147235	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.434	152	62081	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	44560	48.722	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	97.440%
35) Dibromofluoromethane	7.722	113	47866	51.834	ug/l	-0.01
Spiked Amount	50.000	Range	54 - 147	Recovery	=	103.660%
50) Toluene-d8	10.185	98	190821	50.957	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	101.920%
62) 4-Bromofluorobenzene	12.489	95	58153	47.814	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	95.620%
Target Compounds						Qvalue
20) Methylene Chloride	4.729	84	12032	8.360	ug/l	# 76

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY112223\
Data File : VY016450.D
Acq On : 22 Nov 2023 09:37
Operator : SY/MD
Sample : VY1122SBL01
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VY1122SBL01

Quant Time: Nov 23 04:02:17 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y110823S.M
Quant Title : SW846 8260
QLast Update : Thu Nov 09 00:45:52 2023
Response via : Initial Calibration

