

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY112223\
 Data File : VY016454.D
 Acq On : 22 Nov 2023 12:03
 Operator : SY/MD
 Sample : 05503-01
 Misc : 5.01g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 358

Quant Time: Nov 23 04:04:58 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y110823S.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 09 00:45:52 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	108178	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	185491	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.502	117	182067	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.434	152	84966	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	62920	63.639	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	127.280%
35) Dibromofluoromethane	7.728	113	58637	55.648	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	111.300%
50) Toluene-d8	10.191	98	216705	50.715	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	101.420%
62) 4-Bromofluorobenzene	12.489	95	76697	55.265	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	110.520%
Target Compounds						Qvalue
20) Methylene Chloride	4.729	84	15906	11.056	ug/l #	82
95) Naphthalene	15.245	128	261942	97.874	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY112223\
Data File : VY016454.D
Acq On : 22 Nov 2023 12:03
Operator : SY/MD
Sample : 05503-01
Misc : 5.01g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
358

Quant Time: Nov 23 04:04:58 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y110823S.M
Quant Title : SW846 8260
QLast Update : Thu Nov 09 00:45:52 2023
Response via : Initial Calibration

