

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY112223\
 Data File : VY016458.D
 Acq On : 22 Nov 2023 13:36
 Operator : SY/MD
 Sample : BLK
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 BLK

Quant Time: Nov 23 04:05:53 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y110823S.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 09 00:45:52 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.801	168	104425	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.703	114	183755	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.502	117	179457	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.434	152	83142	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.155	65	60974	63.887	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 127.780%	
35) Dibromofluoromethane	7.728	113	57280	54.874	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 109.740%	
50) Toluene-d8	10.191	98	213127	50.349	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	= 100.700%	
62) 4-Bromofluorobenzene	12.489	95	75994	55.276	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	= 110.560%	
Target Compounds						
20) Methylene Chloride	4.729	84	16606	12.263	ug/l	Qvalue # 80

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY112223\
Data File : VY016458.D
Acq On : 22 Nov 2023 13:36
Operator : SY/MD
Sample : BLK
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
BLK

Quant Time: Nov 23 04:05:53 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y110823S.M
Quant Title : SW846 8260
QLast Update : Thu Nov 09 00:45:52 2023
Response via : Initial Calibration

