

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY112223\
 Data File : VY016462.D
 Acq On : 22 Nov 2023 15:10
 Operator : SY/MD
 Sample : 05294-05
 Misc : 5.37g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 WT-5

Quant Time: Nov 23 04:23:03 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y110823S.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 09 00:45:52 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	98558	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	172616	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.496	117	165252	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.434	152	76029	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.143	65	56466	62.686	ug/l	-0.01
Spiked Amount	50.000	Range	50 - 163	Recovery	=	125.380%
35) Dibromofluoromethane	7.722	113	54524	55.605	ug/l	-0.01
Spiked Amount	50.000	Range	54 - 147	Recovery	=	111.200%
50) Toluene-d8	10.185	98	202303	50.876	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	101.760%
62) 4-Bromofluorobenzene	12.489	95	69572	53.870	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	107.740%
Target Compounds						
						Qvalue
20) Methylene Chloride	4.716	84	23044	19.789	ug/l #	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY112223\
Data File : VY016462.D
Acq On : 22 Nov 2023 15:10
Operator : SY/MD
Sample : 05294-05
Misc : 5.37g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
WT-5

Quant Time: Nov 23 04:23:03 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y110823S.M
Quant Title : SW846 8260
QLast Update : Thu Nov 09 00:45:52 2023
Response via : Initial Calibration

