

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY112223\
 Data File : VY016463.D
 Acq On : 22 Nov 2023 15:33
 Operator : SY/MD
 Sample : 05504-01
 Misc : 5.08g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 281

Quant Time: Nov 23 04:07:17 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y110823S.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 09 00:45:52 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	92071	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	159863	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.496	117	158170	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.434	152	69329	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	53279	63.315	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 126.620%	
35) Dibromofluoromethane	7.722	113	51792	57.032	ug/l	-0.01
Spiked Amount	50.000	Range	54 - 147	Recovery	= 114.060%	
50) Toluene-d8	10.185	98	185990	50.505	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	= 101.000%	
62) 4-Bromofluorobenzene	12.489	95	64421	53.861	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	= 107.720%	
Target Compounds						Qvalue
16) Acetone	3.960	43	2191	15.741	ug/l #	75
20) Methylene Chloride	4.716	84	16563	14.363	ug/l #	81

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY112223\
Data File : VY016463.D
Acq On : 22 Nov 2023 15:33
Operator : SY/MD
Sample : 05504-01
Misc : 5.08g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
281

Quant Time: Nov 23 04:07:17 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y110823S.M
Quant Title : SW846 8260
QLast Update : Thu Nov 09 00:45:52 2023
Response via : Initial Calibration

