

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY112223\
 Data File : VY016466.D
 Acq On : 22 Nov 2023 16:44
 Operator : SY/MD
 Sample : 05505-01
 Misc : 4.99g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 TP-A

Quant Time: Nov 23 04:08:21 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y110823S.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 09 00:45:52 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	96825	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	170941	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.495	117	169125	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.434	152	77105	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	58808	66.454	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	132.900%
35) Dibromofluoromethane	7.722	113	54822	56.456	ug/l	-0.01
Spiked Amount	50.000	Range	54 - 147	Recovery	=	112.920%
50) Toluene-d8	10.185	98	202788	51.498	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	103.000%
62) 4-Bromofluorobenzene	12.489	95	70087	54.801	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	109.600%
Target Compounds						
					Qvalue	
20) Methylene Chloride	4.722	84	36778	34.486	ug/l #	83

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY112223\
Data File : VY016466.D
Acq On : 22 Nov 2023 16:44
Operator : SY/MD
Sample : 05505-01
Misc : 4.99g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
TP-A

Quant Time: Nov 23 04:08:21 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y110823S.M
Quant Title : SW846 8260
QLast Update : Thu Nov 09 00:45:52 2023
Response via : Initial Calibration

