

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY112223\
 Data File : VY016471.D
 Acq On : 22 Nov 2023 18:41
 Operator : SY/MD
 Sample : 05540-01
 Misc : 5.02g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 TR-06-11222023

Quant Time: Nov 23 04:09:54 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y110823S.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 09 00:45:52 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	24248	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	37907	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.496	117	34006	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.434	152	13520	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	12393	55.921	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	111.840%
35) Dibromofluoromethane	7.728	113	12874	59.786	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	119.580%
50) Toluene-d8	10.191	98	42649	48.840	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	97.680%
62) 4-Bromofluorobenzene	12.489	95	12188	42.974	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	85.940%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY112223\
Data File : VY016471.D
Acq On : 22 Nov 2023 18:41
Operator : SY/MD
Sample : 05540-01
Misc : 5.02g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
TR-06-11222023

Quant Time: Nov 23 04:09:54 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y110823S.M
Quant Title : SW846 8260
QLast Update : Thu Nov 09 00:45:52 2023
Response via : Initial Calibration

