

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY112223\
 Data File : VY016449.D
 Acq On : 22 Nov 2023 08:46
 Operator : SY/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 LabSampleId :
 VSTDCCC050

Quant Time: Nov 23 04:01:28 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y110823S.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 09 00:45:52 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	91	0.00
2 T	Dichlorodifluoromethane	0.490	0.485	1.0	96	0.00
3 P	Chloromethane	0.573	0.506	11.7	85	0.00
4 C	Vinyl Chloride	0.617	0.562	8.9#	87	0.00
5 T	Bromomethane	0.405	0.382	5.7	90	0.00
6 T	Chloroethane	0.393	0.371	5.6	90	0.00
7 T	Trichlorofluoromethane	0.897	0.972	-8.4	99	0.00
8 T	Diethyl Ether	0.268	0.285	-6.3	97	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.540	0.556	-3.0	97	0.00
10 T	Methyl Iodide	0.584	0.662	-13.4	101	0.00
11 T	Tert butyl alcohol	0.036	0.035	2.8	92	0.00
12 CM	1,1-Dichloroethene	0.505	0.531	-5.1#	98	0.00
13 T	Acrolein	0.043	0.035	18.6	73	0.00
14 T	Allyl chloride	0.717	0.655	8.6	82	0.00
15 T	Acrylonitrile	0.106	0.107	-0.9	89	0.00
16 T	Acetone	0.076	0.093	-22.4	116	0.00
17 T	Carbon Disulfide	1.525	1.560	-2.3	95	0.00
18 T	Methyl Acetate	0.402	0.388	3.5	87	0.00
19 T	Methyl tert-butyl Ether	1.223	1.267	-3.6	93	0.00
20 T	Methylene Chloride	0.610	0.589	3.4	97	0.00
21 T	trans-1,2-Dichloroethene	0.582	0.602	-3.4	97	0.00
22 T	Diisopropyl ether	1.545	1.404	9.1	81	0.00
23 T	Vinyl Acetate	0.749	0.691	7.7	81	0.00
24 P	1,1-Dichloroethane	0.953	0.931	2.3	90	-0.01
25 T	2-Butanone	0.132	0.132	0.0	91	0.00
26 T	2,2-Dichloropropane	0.868	0.903	-4.0	95	0.00
27 T	cis-1,2-Dichloroethene	0.627	0.655	-4.5	96	0.00
28 T	Bromochloromethane	0.353	0.370	-4.8	96	0.00
29 T	Tetrahydrofuran	0.088	0.082	6.8	82	0.00
30 C	Chloroform	0.992	1.019	-2.7#	95	0.00
31 T	Cyclohexane	0.900	0.816	9.3	87	0.00
32 T	1,1,1-Trichloroethane	0.910	0.971	-6.7	98	0.00
33 S	1,2-Dichloroethane-d4	0.457	0.456	0.2	93	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	88	0.00
35 S	Dibromofluoromethane	0.284	0.321	-13.0	100	0.00
36 T	1,1-Dichloropropene	0.493	0.510	-3.4	91	0.00
37 T	Ethyl Acetate	0.199	0.187	6.0	82	0.00
38 T	Carbon Tetrachloride	0.512	0.586	-14.5	101	0.00
39 T	Methylcyclohexane	0.615	0.653	-6.2	93	0.00
40 TM	Benzene	1.386	1.467	-5.8	93	0.00
41 T	Methacrylonitrile	0.100	0.103	-3.0	92	0.00
42 TM	1,2-Dichloroethane	0.359	0.383	-6.7	92	0.00
43 T	Isopropyl Acetate	0.385	0.366	4.9	81	0.00
44 TM	Trichloroethene	0.398	0.443	-11.3	98	0.00
45 C	1,2-Dichloropropane	0.333	0.336	-0.9#	87	0.00
46 T	Dibromomethane	0.180	0.196	-8.9	93	0.00
47 T	Bromodichloromethane	0.464	0.499	-7.5	94	0.00
48 T	Methyl methacrylate	0.213	0.209	1.9	82	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY112223\
 Data File : VY016449.D
 Acq On : 22 Nov 2023 08:46
 Operator : SY/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 LabSampleId :
 VSTDCCC050

Quant Time: Nov 23 04:01:28 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y110823S.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 09 00:45:52 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.002	0.002	0.0	97	0.00
50 S	Toluene-d8	1.152	1.217	-5.6	94	0.00
51 T	4-Methyl-2-Pentanone	0.195	0.188	3.6	82	0.00
52 CM	Toluene	0.876	0.943	-7.6#	94	0.00
53 T	t-1,3-Dichloropropene	0.449	0.487	-8.5	93	0.00
54 T	cis-1,3-Dichloropropene	0.537	0.577	-7.4	92	0.00
55 T	1,1,2-Trichloroethane	0.249	0.273	-9.6	94	0.00
56 T	Ethyl methacrylate	0.235	0.262	-11.5	92	0.00
57 T	1,3-Dichloropropane	0.424	0.458	-8.0	92	0.00
58 T	2-Chloroethyl Vinyl ether	0.159	0.185	-16.4	92	0.00
59 T	2-Hexanone	0.145	0.139	4.1	86	0.00
60 T	Dibromochloromethane	0.311	0.363	-16.7	99	0.00
61 T	1,2-Dibromoethane	0.237	0.269	-13.5	97	0.00
62 S	4-Bromofluorobenzene	0.374	0.416	-11.2	99	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	89	0.00
64 T	Tetrachloroethene	0.449	0.503	-12.0	99	0.00
65 PM	Chlorobenzene	1.068	1.171	-9.6	97	0.00
66 T	1,1,1,2-Tetrachloroethane	0.394	0.439	-11.4	98	0.00
67 C	Ethyl Benzene	1.920	2.049	-6.7#	93	0.00
68 T	m/p-Xylenes	0.728	0.799	-9.8	96	0.00
69 T	o-Xylene	0.676	0.747	-10.5	95	0.00
70 T	Styrene	0.968	1.099	-13.5	97	0.00
71 P	Bromoform	0.204	0.241	-18.1	101	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	93	0.00
73 T	Isopropylbenzene	4.049	4.237	-4.6	97	0.00
74 T	N-amyl acetate	0.813	0.742	8.7	80	0.00
75 P	1,1,2,2-Tetrachloroethane	0.650	0.667	-2.6	95	0.00
76 T	1,2,3-Trichloropropane	0.480	0.506	-5.4	104	0.00
77 T	Bromobenzene	0.915	0.986	-7.8	99	0.00
78 T	n-propylbenzene	4.785	4.873	-1.8	94	0.00
79 T	2-Chlorotoluene	2.637	2.704	-2.5	94	0.00
80 T	1,3,5-Trimethylbenzene	3.264	3.448	-5.6	97	0.00
81 T	trans-1,4-Dichloro-2-butene	0.221	0.227	-2.7	94	0.00
82 T	4-Chlorotoluene	2.725	2.822	-3.6	96	0.00
83 T	tert-Butylbenzene	2.875	3.025	-5.2	96	0.00
84 T	1,2,4-Trimethylbenzene	3.214	3.407	-6.0	97	0.00
85 T	sec-Butylbenzene	4.108	4.266	-3.8	96	0.00
86 T	p-Isopropyltoluene	3.537	3.799	-7.4	98	0.00
87 T	1,3-Dichlorobenzene	1.761	1.882	-6.9	100	0.00
88 T	1,4-Dichlorobenzene	1.721	1.849	-7.4	100	0.00
89 T	n-Butylbenzene	3.267	3.408	-4.3	95	0.00
90 T	Hexachloroethane	0.682	0.725	-6.3	99	0.00
91 T	1,2-Dichlorobenzene	1.509	1.646	-9.1	102	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.103	0.104	-1.0	97	0.00
93 T	1,2,4-Trichlorobenzene	0.881	1.048	-19.0	110	0.00
94 T	Hexachlorobutadiene	0.504	0.599	-18.8	112	0.00
95 T	Naphthalene	1.575	1.929	-22.5	112	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY112223\
Data File : VY016449.D
Acq On : 22 Nov 2023 08:46
Operator : SY/MD
Sample : VSTDCCC050
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_Y
LabSampleId :
VSTDCCC050

Quant Time: Nov 23 04:01:28 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y110823S.M
Quant Title : SW846 8260
QLast Update : Thu Nov 09 00:45:52 2023
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.728	0.892	-22.5	117	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6