

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY112224\
 Data File : VY020409.D
 Acq On : 22 Nov 2024 13:39
 Operator : SY/MD
 Sample : P4948-01
 Misc : 5.08g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 337

Quant Time: Nov 23 00:00:06 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y111924S.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 20 04:38:24 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.719	168	144691	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.622	114	276397	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.420	117	259166	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.352	152	100534	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.067	65	90740	54.586	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	109.180%
35) Dibromofluoromethane	7.640	113	89062	50.251	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	100.500%
50) Toluene-d8	10.109	98	334819	47.957	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	95.920%
62) 4-Bromofluorobenzene	12.407	95	111096	49.500	ug/l	0.00
Spiked Amount	50.000	Range	29 - 146	Recovery	=	99.000%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY112224\
Data File : VY020409.D
Acq On : 22 Nov 2024 13:39
Operator : SY/MD
Sample : P4948-01
Misc : 5.08g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
337

Quant Time: Nov 23 00:00:06 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y111924S.M
Quant Title : SW846 8260
QLast Update : Wed Nov 20 04:38:24 2024
Response via : Initial Calibration

