

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY112224\
 Data File : VY020413.D
 Acq On : 22 Nov 2024 15:13
 Operator : SY/MD
 Sample : P4946-01
 Misc : 5.05g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 BW-1

Quant Time: Nov 23 00:01:55 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y111924S.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 20 04:38:24 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.713	168	137287	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.622	114	258536	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.420	117	238609	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.353	152	94206	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.067	65	82172	52.098	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	104.200%
35) Dibromofluoromethane	7.640	113	80043	48.282	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	96.560%
50) Toluene-d8	10.109	98	310464	47.541	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	95.080%
62) 4-Bromofluorobenzene	12.408	95	103571	49.335	ug/l	0.00
Spiked Amount	50.000	Range	29 - 146	Recovery	=	98.680%
Target Compounds						
51) 4-Methyl-2-Pentanone	9.999	43	1458	1.274	ug/l	Qvalue 89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY112224\
Data File : VY020413.D
Acq On : 22 Nov 2024 15:13
Operator : SY/MD
Sample : P4946-01
Misc : 5.05g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
BW-1

Quant Time: Nov 23 00:01:55 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y111924S.M
Quant Title : SW846 8260
QLast Update : Wed Nov 20 04:38:24 2024
Response via : Initial Calibration

