

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY112224\
 Data File : VY020419.D
 Acq On : 22 Nov 2024 17:33
 Operator : SY/MD
 Sample : P4960-06
 Misc : 5.01g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SW4

Quant Time: Nov 23 00:04:00 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y111924S.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 20 04:38:24 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.719	168	95	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.622	114	323	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.420	117	86	50.000	ug/l	# 0.00
72) 1,4-Dichlorobenzene-d4	13.365	152	21	50.000	ug/l	# 0.02
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.079	65	169	154.841	ug/l	0.02
Spiked Amount	50.000	Range	50 - 163	Recovery	=	309.680%#
35) Dibromofluoromethane	7.634	113	113	54.559	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	109.120%
50) Toluene-d8	10.121	98	280	34.319	ug/l	0.01
Spiked Amount	50.000	Range	58 - 134	Recovery	=	68.640%
62) 4-Bromofluorobenzene	0.000	95	0	0.000	ug/l	
Spiked Amount	50.000	Range	29 - 146	Recovery	=	0.000%#

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY112224\
Data File : VY020419.D
Acq On : 22 Nov 2024 17:33
Operator : SY/MD
Sample : P4960-06
Misc : 5.01g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SW4

Quant Time: Nov 23 00:04:00 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y111924S.M
Quant Title : SW846 8260
QLast Update : Wed Nov 20 04:38:24 2024
Response via : Initial Calibration

