

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY112624\
 Data File : VY020447.D
 Acq On : 26 Nov 2024 13:34
 Operator : SY/MD
 Sample : P4985-07
 Misc : 3.87g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 MH-740-VOC

Quant Time: Nov 27 01:40:46 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y111924S.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 20 04:38:24 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.713	168	117997	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.616	114	221277	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.414	117	208724	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.347	152	81343	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.061	65	72273	53.312	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	106.620%
35) Dibromofluoromethane	7.640	113	70234	49.499	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	99.000%
50) Toluene-d8	10.109	98	268952	48.119	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	96.240%
62) 4-Bromofluorobenzene	12.408	95	90427	50.327	ug/l	0.00
Spiked Amount	50.000	Range	29 - 146	Recovery	=	100.660%
Target Compounds						
16) Acetone	3.879	43	580m	1.785	ug/l	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY112624\
Data File : VY020447.D
Acq On : 26 Nov 2024 13:34
Operator : SY/MD
Sample : P4985-07
Misc : 3.87g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
MH-740-VOC

Quant Time: Nov 27 01:40:46 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y111924S.M
Quant Title : SW846 8260
QLast Update : Wed Nov 20 04:38:24 2024
Response via : Initial Calibration

