

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY112625\
 Data File : VY023820.D
 Acq On : 26 Nov 2025 12:34
 Operator : SY/MD
 Sample : Q3604-04
 Misc : 5.06g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 S-1

Quant Time: Nov 27 01:19:57 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y112625S.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 27 01:14:36 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.713	168	742936	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.616	114	1381435	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.414	117	1169428	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.346	152	458839	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.061	65	418065	57.783	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	115.560%
35) Dibromofluoromethane	7.634	113	433941	53.027	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	106.060%
50) Toluene-d8	10.109	98	1631756	50.201	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	100.400%
62) 4-Bromofluorobenzene	12.408	95	472576	44.186	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	88.380%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY112625\
Data File : VY023820.D
Acq On : 26 Nov 2025 12:34
Operator : SY/MD
Sample : Q3604-04
Misc : 5.06g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
S-1

Quant Time: Nov 27 01:19:57 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y112625S.M
Quant Title : SW846 8260
QLast Update : Thu Nov 27 01:14:36 2025
Response via : Initial Calibration

