

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY112625\
 Data File : VY023826.D
 Acq On : 26 Nov 2025 14:54
 Operator : SY/MD
 Sample : VIBLK
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VIBLK

Quant Time: Nov 27 01:23:11 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y112625S.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 27 01:14:36 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.707	168	749382	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.615	114	1381356	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.414	117	1194676	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.346	152	495364	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.060	65	436549	59.819	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	119.640%
35) Dibromofluoromethane	7.634	113	432409	52.842	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	105.680%
50) Toluene-d8	10.109	98	1625972	50.025	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	100.060%
62) 4-Bromofluorobenzene	12.401	95	487645	45.598	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	91.200%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY112625\
Data File : VY023826.D
Acq On : 26 Nov 2025 14:54
Operator : SY/MD
Sample : VIBLK
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VIBLK

Quant Time: Nov 27 01:23:11 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y112625S.M
Quant Title : SW846 8260
QLast Update : Thu Nov 27 01:14:36 2025
Response via : Initial Calibration

