

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY112724\
 Data File : VY020456.D
 Acq On : 27 Nov 2024 11:01
 Operator : SY/MD
 Sample : VY1127SBL01
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VY1127SBL01

Quant Time: Nov 28 00:57:18 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y111924S.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 20 04:38:24 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.719	168	135718	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.622	114	259964	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.426	117	233221	50.000	ug/l	0.01
72) 1,4-Dichlorobenzene-d4	13.353	152	82357	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.073	65	77468	49.683	ug/l	0.01
Spiked Amount	50.000	Range	50 - 163	Recovery	=	99.360%
35) Dibromofluoromethane	7.646	113	79429	47.649	ug/l	0.01
Spiked Amount	50.000	Range	54 - 147	Recovery	=	95.300%
50) Toluene-d8	10.115	98	311581	47.450	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	94.900%
62) 4-Bromofluorobenzene	12.414	95	95697	45.334	ug/l	0.01
Spiked Amount	50.000	Range	29 - 146	Recovery	=	90.660%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSV0A_Y\Data\VY112724\
Data File : VY020456.D
Acq On : 27 Nov 2024 11:01
Operator : SY/MD
Sample : VY1127SBL01
Misc : 5.00g/5.0mL/MSV0A_Y/SOIL
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSV0A_Y
ClientSampleId :
VY1127SBL01

Quant Time: Nov 28 00:57:18 2024
Quant Method : Z:\voasrv\HPCHEM1\MSV0A_Y\methods\82Y111924S.M
Quant Title : SW846 8260
QLast Update : Wed Nov 20 04:38:24 2024
Response via : Initial Calibration

