

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY112724\
 Data File : VY020460.D
 Acq On : 27 Nov 2024 12:57
 Operator : SY/MD
 Sample : P5019-01
 Misc : 5.05g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 EO-02-11262024

Quant Time: Nov 28 00:59:36 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y111924S.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 20 04:38:24 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.726	168	492	50.000	ug/l	# 0.01
34) 1,4-Difluorobenzene	8.616	114	1034	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.426	117	461	50.000	ug/l	# 0.01
72) 1,4-Dichlorobenzene-d4	13.340	152	75	50.000	ug/l	# 0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.061	65	485	85.803	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	171.600%#
35) Dibromofluoromethane	7.640	113	292	44.040	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	88.080%
50) Toluene-d8	10.109	98	977	37.407	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	74.820%
62) 4-Bromofluorobenzene	12.408	95	79	9.409	ug/l	0.00
Spiked Amount	50.000	Range	29 - 146	Recovery	=	18.820%#

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSV0A_Y\Data\VY112724\
Data File : VY020460.D
Acq On : 27 Nov 2024 12:57
Operator : SY/MD
Sample : P5019-01
Misc : 5.05g/5.0mL/MSV0A_Y/SOIL/A
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSV0A_Y
ClientSampleId :
EO-02-11262024

Quant Time: Nov 28 00:59:36 2024
Quant Method : Z:\voasrv\HPCHEM1\MSV0A_Y\methods\82Y111924S.M
Quant Title : SW846 8260
QLast Update : Wed Nov 20 04:38:24 2024
Response via : Initial Calibration

