

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY112724\
 Data File : VY020461.D
 Acq On : 27 Nov 2024 13:20
 Operator : SY/MD
 Sample : P5005-01
 Misc : 5.07g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 STOCK-PILE

Quant Time: Nov 28 00:59:55 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y111924S.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 20 04:38:24 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.713	168	1696	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.622	114	2829	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.420	117	1817	50.000	ug/l	# 0.00
72) 1,4-Dichlorobenzene-d4	13.352	152	321	50.000	ug/l	# 0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.073	65	1165	59.789	ug/l	0.01
Spiked Amount	50.000	Range	50 - 163	Recovery	=	119.580%
35) Dibromofluoromethane	7.634	113	957	52.755	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	105.520%
50) Toluene-d8	10.115	98	2856	39.967	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	79.940%
62) 4-Bromofluorobenzene	12.408	95	323	14.061	ug/l	0.00
Spiked Amount	50.000	Range	29 - 146	Recovery	=	28.120%#

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSV0A_Y\Data\VY112724\
Data File : VY020461.D
Acq On : 27 Nov 2024 13:20
Operator : SY/MD
Sample : P5005-01
Misc : 5.07g/5.0mL/MSV0A_Y/SOIL/A
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSV0A_Y
ClientSampleId :
STOCK-PILE

Quant Time: Nov 28 00:59:55 2024
Quant Method : Z:\voasrv\HPCHEM1\MSV0A_Y\methods\82Y111924S.M
Quant Title : SW846 8260
QLast Update : Wed Nov 20 04:38:24 2024
Response via : Initial Calibration

