

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY112724\
 Data File : VY020465.D
 Acq On : 27 Nov 2024 17:01
 Operator : SY/MD
 Sample : P5026-04
 Misc : 6.28g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SOIL-1-HAM-GRAB

Quant Time: Nov 28 01:02:03 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y111924S.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 20 04:38:24 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.713	168	134136	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.622	114	262414	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.420	117	254042	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.353	152	103875	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.067	65	83756	54.349	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	108.700%
35) Dibromofluoromethane	7.640	113	64766	38.490	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	76.980%
50) Toluene-d8	10.109	98	321173	48.454	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	96.900%
62) 4-Bromofluorobenzene	12.408	95	112056	52.588	ug/l	0.00
Spiked Amount	50.000	Range	29 - 146	Recovery	=	105.180%
Target Compounds						
					Qvalue	
17) Carbon Disulfide	4.110	76	12978	2.633	ug/l	100
52) Toluene	10.170	92	5702	1.239	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSV0A_Y\Data\VY112724\
Data File : VY020465.D
Acq On : 27 Nov 2024 17:01
Operator : SY/MD
Sample : P5026-04
Misc : 6.28g/5.0mL/MSV0A_Y/SOIL/A
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SOIL-1-HAM-GRAB

Quant Time: Nov 28 01:02:03 2024
Quant Method : Z:\voasrv\HPCHEM1\MSV0A_Y\methods\82Y111924S.M
Quant Title : SW846 8260
QLast Update : Wed Nov 20 04:38:24 2024
Response via : Initial Calibration

