

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY112724\
 Data File : VY020467.D
 Acq On : 27 Nov 2024 17:48
 Operator : SY/MD
 Sample : P5045-01
 Misc : 5.05g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SU-04-11222024

Quant Time: Nov 28 01:02:58 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y111924S.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 20 04:38:24 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.713	168	23489	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.616	114	35910	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.420	117	26049	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.353	152	7112	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.061	65	8939	33.124	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	66.240%
35) Dibromofluoromethane	7.646	113	10366	45.018	ug/l	0.01
Spiked Amount	50.000	Range	54 - 147	Recovery	=	90.040%
50) Toluene-d8	10.109	98	40696	44.866	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	89.740%
62) 4-Bromofluorobenzene	12.408	95	10175	34.895	ug/l	0.00
Spiked Amount	50.000	Range	29 - 146	Recovery	=	69.780%

Target Compounds Qvalue

 (#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSV0A_Y\Data\VY112724\
Data File : VY020467.D
Acq On : 27 Nov 2024 17:48
Operator : SY/MD
Sample : P5045-01
Misc : 5.05g/5.0mL/MSV0A_Y/SOIL/A
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSV0A_Y
ClientSampleId :
SU-04-11222024

Quant Time: Nov 28 01:02:58 2024
Quant Method : Z:\voasrv\HPCHEM1\MSV0A_Y\methods\82Y111924S.M
Quant Title : SW846 8260
QLast Update : Wed Nov 20 04:38:24 2024
Response via : Initial Calibration

