

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY112724\
 Data File : VY020469.D
 Acq On : 27 Nov 2024 18:34
 Operator : SY/MD
 Sample : P5046-01
 Misc : 5.05g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 HD-01-11272024

Quant Time: Nov 28 01:03:38 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y111924S.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 20 04:38:24 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.713	168	452	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.622	114	960	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.420	117	406	50.000	ug/l	# 0.00
72) 1,4-Dichlorobenzene-d4	13.346	152	124	50.000	ug/l	# 0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.061	65	437	84.152	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	168.300%#
35) Dibromofluoromethane	7.622	113	228	37.038	ug/l	-0.01
Spiked Amount	50.000	Range	54 - 147	Recovery	=	74.080%
50) Toluene-d8	10.109	98	721	29.733	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	59.460%
62) 4-Bromofluorobenzene	0.000	95	0	0.000	ug/l	
Spiked Amount	50.000	Range	29 - 146	Recovery	=	0.000%#

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSV0A_Y\Data\VY112724\
Data File : VY020469.D
Acq On : 27 Nov 2024 18:34
Operator : SY/MD
Sample : P5046-01
Misc : 5.05g/5.0mL/MSV0A_Y/SOIL/A
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSV0A_Y
ClientSampleId :
HD-01-11272024

Quant Time: Nov 28 01:03:38 2024
Quant Method : Z:\voasrv\HPCHEM1\MSV0A_Y\methods\82Y111924S.M
Quant Title : SW846 8260
QLast Update : Wed Nov 20 04:38:24 2024
Response via : Initial Calibration

