

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY112922\
 Data File : VY011615.D
 Acq On : 29 Nov 2022 12:26
 Operator : KP/MD
 Sample : N5754-01
 Misc : 5.03g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 PSC-124019

Quant Time: Nov 29 12:52:18 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y111822S.M
 Quant Title : SW846 8260
 QLast Update : Sat Nov 19 22:37:14 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	128115	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	237784	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	229524	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	98265	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.143	65	95388	73.326	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	146.660%
35) Dibromofluoromethane	7.722	113	80981	55.323	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	110.640%
50) Toluene-d8	10.179	98	287197	53.992	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	107.980%
62) 4-Bromofluorobenzene	12.483	95	93652	47.619	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	95.240%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY112922\
Data File : VY011615.D
Acq On : 29 Nov 2022 12:26
Operator : KP/MD
Sample : N5754-01
Misc : 5.03g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
PSC-124019

Quant Time: Nov 29 12:52:18 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y111822S.M
Quant Title : SW846 8260
QLast Update : Sat Nov 19 22:37:14 2022
Response via : Initial Calibration

