

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY112922\
 Data File : VY011621.D
 Acq On : 29 Nov 2022 14:47
 Operator : KP/MD
 Sample : N5789-01
 Misc : 5.01g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 TFI-359

Quant Time: Nov 30 04:43:23 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y111822S.M
 Quant Title : SW846 8260
 QLast Update : Sat Nov 19 22:37:14 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	3214	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	6513	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.496	117	7447	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	2145	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.142	65	3929	120.393	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	240.780%#
35) Dibromofluoromethane	7.716	113	1539	38.385	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	76.780%
50) Toluene-d8	10.185	98	7930	54.428	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	108.860%
62) 4-Bromofluorobenzene	12.483	95	2723	50.549	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	101.100%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY112922\
Data File : VY011621.D
Acq On : 29 Nov 2022 14:47
Operator : KP/MD
Sample : N5789-01
Misc : 5.01g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
TFI-359

Quant Time: Nov 30 04:43:23 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y111822S.M
Quant Title : SW846 8260
QLast Update : Sat Nov 19 22:37:14 2022
Response via : Initial Calibration

