

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY120125\
 Data File : VY023830.D
 Acq On : 01 Dec 2025 09:13
 Operator : SY/MD
 Sample : VY1201SBL01
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VY1201SBL01

Quant Time: Dec 02 03:12:33 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y112625S.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 27 01:14:36 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.713	168	719047	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.622	114	1221329	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.420	117	995220	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.346	152	399677	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.067	65	408084	58.278	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	116.560%
35) Dibromofluoromethane	7.640	113	368921	50.991	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	101.980%
50) Toluene-d8	10.109	98	1423543	49.536	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	99.080%
62) 4-Bromofluorobenzene	12.408	95	399287	42.228	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	84.460%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY120125\
Data File : VY023830.D
Acq On : 01 Dec 2025 09:13
Operator : SY/MD
Sample : VY1201SBL01
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VY1201SBL01

Quant Time: Dec 02 03:12:33 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y112625S.M
Quant Title : SW846 8260
QLast Update : Thu Nov 27 01:14:36 2025
Response via : Initial Calibration

