

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY120125\
 Data File : VY023839.D
 Acq On : 01 Dec 2025 13:36
 Operator : SY/MD
 Sample : Q3741-04
 Misc : 5.64g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 B50-SP9-112425

Quant Time: Dec 02 03:16:48 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y112625S.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 27 01:14:36 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.713	168	491424	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.616	114	929861	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.420	117	778289	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.347	152	274066	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.061	65	326284	68.179	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	136.360%
35) Dibromofluoromethane	7.634	113	135756	24.645	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	49.300%#
50) Toluene-d8	10.109	98	1111835	50.817	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	101.640%
62) 4-Bromofluorobenzene	12.408	95	291322	40.467	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	80.940%
Target Compounds						
16) Acetone	3.879	43	17509	12.698	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY120125\
Data File : VY023839.D
Acq On : 01 Dec 2025 13:36
Operator : SY/MD
Sample : Q3741-04
Misc : 5.64g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
B50-SP9-112425

Quant Time: Dec 02 03:16:48 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y112625S.M
Quant Title : SW846 8260
QLast Update : Thu Nov 27 01:14:36 2025
Response via : Initial Calibration

