

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY120125\
 Data File : VY023843.D
 Acq On : 01 Dec 2025 15:10
 Operator : SY/MD
 Sample : Q3742-01
 Misc : 2.62g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SB-1125-23

Quant Time: Dec 02 03:18:25 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y112625S.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 27 01:14:36 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.713	168	656114	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.616	114	1207708	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.414	117	1048441	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.346	152	433339	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.061	65	390837	61.168	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	122.340%
35) Dibromofluoromethane	7.634	113	386244	53.987	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	107.980%
50) Toluene-d8	10.109	98	1445760	50.877	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	101.760%
62) 4-Bromofluorobenzene	12.408	95	425798	45.539	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	91.080%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY120125\
Data File : VY023843.D
Acq On : 01 Dec 2025 15:10
Operator : SY/MD
Sample : Q3742-01
Misc : 2.62g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SB-1125-23

Quant Time: Dec 02 03:18:25 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y112625S.M
Quant Title : SW846 8260
QLast Update : Thu Nov 27 01:14:36 2025
Response via : Initial Calibration

