

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY120125\
 Data File : VY023845.D
 Acq On : 01 Dec 2025 15:57
 Operator : SY/MD
 Sample : Q3742-03
 Misc : 2.50g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SB-1125-8

Quant Time: Dec 02 03:19:10 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y112625S.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 27 01:14:36 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.713	168	608834	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.616	114	1113703	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.414	117	955695	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.346	152	376385	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.061	65	354950	59.866	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	119.740%
35) Dibromofluoromethane	7.634	113	355083	53.821	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	107.640%
50) Toluene-d8	10.109	98	1331655	50.817	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	101.640%
62) 4-Bromofluorobenzene	12.408	95	383301	44.454	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	88.900%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY120125\
Data File : VY023845.D
Acq On : 01 Dec 2025 15:57
Operator : SY/MD
Sample : Q3742-03
Misc : 2.50g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SB-1125-8

Quant Time: Dec 02 03:19:10 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y112625S.M
Quant Title : SW846 8260
QLast Update : Thu Nov 27 01:14:36 2025
Response via : Initial Calibration

