

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY120925\
 Data File : VY023906.D
 Acq On : 09 Dec 2025 14:19
 Operator : SY/MD
 Sample : Q3790-03RE
 Misc : 5.43g/5.0mL/MSVOA_Y/SOIL/B
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 304 FURMAN SOILRE

Quant Time: Dec 10 02:27:53 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y120325S.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 04 03:49:13 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.725	168	10619	50.000	ug/l	# 0.01
34) 1,4-Difluorobenzene	8.634	114	15625	50.000	ug/l	0.02
63) Chlorobenzene-d5	11.426	117	7773	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.353	152	1424	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.079	65	5210	49.045	ug/l	0.01
Spiked Amount	50.000	Range	50 - 163	Recovery	=	98.100%
35) Dibromofluoromethane	7.652	113	4396	43.866	ug/l	0.01
Spiked Amount	50.000	Range	54 - 147	Recovery	=	87.740%
50) Toluene-d8	10.121	98	14270	36.592	ug/l	0.01
Spiked Amount	50.000	Range	58 - 134	Recovery	=	73.180%
62) 4-Bromofluorobenzene	12.414	95	1644	12.360	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	24.720%#

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY120925\
Data File : VY023906.D
Acq On : 09 Dec 2025 14:19
Operator : SY/MD
Sample : Q3790-03RE
Misc : 5.43g/5.0mL/MSVOA_Y/SOIL/B
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
304 FURMAN SOILRE

Quant Time: Dec 10 02:27:53 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y120325S.M
Quant Title : SW846 8260
QLast Update : Thu Dec 04 03:49:13 2025
Response via : Initial Calibration

