

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY121124\
 Data File : VY020584.D
 Acq On : 11 Dec 2024 13:37
 Operator : SY/MD
 Sample : P5219-03RE
 Misc : 5.06g/5.0mL/MSVOA_Y/SOIL/B
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 MR-CAM-012RE

Quant Time: Dec 12 00:05:21 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y120224S.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 03 01:39:23 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.707	168	109743	50.000	ug/l	-0.01
34) 1,4-Difluorobenzene	8.609	114	178393	50.000	ug/l	-0.01
63) Chlorobenzene-d5	11.414	117	106734	50.000	ug/l	-0.01
72) 1,4-Dichlorobenzene-d4	13.346	152	14371	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.061	65	62205	55.734	ug/l	-0.01
Spiked Amount	50.000	Range	50 - 163	Recovery	= 111.460%	
35) Dibromofluoromethane	7.634	113	58288	51.901	ug/l	-0.01
Spiked Amount	50.000	Range	54 - 147	Recovery	= 103.800%	
50) Toluene-d8	10.103	98	189689	44.357	ug/l	-0.01
Spiked Amount	50.000	Range	58 - 134	Recovery	= 88.720%	
62) 4-Bromofluorobenzene	12.401	95	32889	21.989	ug/l	-0.01
Spiked Amount	50.000	Range	29 - 146	Recovery	= 43.980%	
Target Compounds						
					Qvalue	
16) Acetone	3.860	43	9339	33.652	ug/l	90
17) Carbon Disulfide	4.110	76	4277	1.312	ug/l	98
20) Methylene Chloride	4.610	84	1309	1.109	ug/l	87
84) 1,2,4-Trimethylbenzene	13.041	105	3910	3.610	ug/l	97
86) p-Isopropyltoluene	13.291	119	1797	1.506	ug/l #	71
96) 1,2,3-Trichlorobenzene	15.328	180	778	3.641	ug/l	85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY121124\
Data File : VY020584.D
Acq On : 11 Dec 2024 13:37
Operator : SY/MD
Sample : P5219-03RE
Misc : 5.06g/5.0mL/MSVOA_Y/SOIL/B
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
MR-CAM-012RE

Quant Time: Dec 12 00:05:21 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y120224S.M
Quant Title : SW846 8260
QLast Update : Tue Dec 03 01:39:23 2024
Response via : Initial Calibration

