



Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY121124\
Data File : VY020579.D
Acq On : 11 Dec 2024 09:12
Operator : SY/MD
Sample : BFB
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y120224S.M
Title : SW846 8260
Last Update : Tue Dec 03 01:39:23 2024



AutoFind: Scans 1752, 1753, 1754; Background Corrected with Scan 1744

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	19.7	26005	PASS
75	95	30	60	54.4	71699	PASS
95	95	100	100	100.0	131899	PASS
96	95	5	9	6.8	8944	PASS
173	174	0.00	2	0.6	626	PASS
174	95	50	100	81.5	107488	PASS
175	174	5	9	7.8	8384	PASS
176	174	95	101	98.1	105496	PASS
177	176	5	9	6.9	7274	PASS