

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY121824\
 Data File : VY020635.D
 Acq On : 18 Dec 2024 14:54
 Operator : SY/MD
 Sample : P5299-04
 Misc : 6.77g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SB-02

Quant Time: Dec 19 02:48:56 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y121724S.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 18 05:27:13 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.707	168	157761	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.610	114	310451	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.414	117	269674	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.347	152	117196	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.061	65	101134	58.533	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	117.060%
35) Dibromofluoromethane	7.634	113	88259	40.338	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	80.680%
50) Toluene-d8	10.103	98	372749	48.673	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	97.340%
62) 4-Bromofluorobenzene	12.402	95	122857	41.011	ug/l	0.00
Spiked Amount	50.000	Range	29 - 146	Recovery	=	82.020%
Target Compounds						
16) Acetone	3.867	43	13742	34.486	ug/l	96
25) 2-Butanone	6.896	43	5232	8.148	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY121824\
Data File : VY020635.D
Acq On : 18 Dec 2024 14:54
Operator : SY/MD
Sample : P5299-04
Misc : 6.77g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SB-02

Quant Time: Dec 19 02:48:56 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y121724S.M
Quant Title : SW846 8260
QLast Update : Wed Dec 18 05:27:13 2024
Response via : Initial Calibration

