

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY121824\
 Data File : VY020637.D
 Acq On : 18 Dec 2024 15:41
 Operator : SY/MD
 Sample : P5277-02
 Misc : 5.08g/5.0mL/MSVOA_Y/SOIL/B
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 COMP-2A

Quant Time: Dec 19 02:49:45 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y121724S.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 18 05:27:13 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.720	168	1353	50.000	ug/l	# 0.01
34) 1,4-Difluorobenzene	8.610	114	2390	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.414	117	1242	50.000	ug/l	# 0.00
72) 1,4-Dichlorobenzene-d4	13.347	152	261	50.000	ug/l	# 0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.067	65	1278	86.245	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	172.500%#
35) Dibromofluoromethane	7.628	113	897	53.253	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	106.500%
50) Toluene-d8	10.103	98	2424	41.115	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	82.220%
62) 4-Bromofluorobenzene	12.402	95	267	11.577	ug/l	0.00
Spiked Amount	50.000	Range	29 - 146	Recovery	=	23.160%#

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY121824\
Data File : VY020637.D
Acq On : 18 Dec 2024 15:41
Operator : SY/MD
Sample : P5277-02
Misc : 5.08g/5.0mL/MSVOA_Y/SOIL/B
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
COMP-2A

Quant Time: Dec 19 02:49:45 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y121724S.M
Quant Title : SW846 8260
QLast Update : Wed Dec 18 05:27:13 2024
Response via : Initial Calibration

