

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY121824\
 Data File : VY020642.D
 Acq On : 18 Dec 2024 17:38
 Operator : SY/MD
 Sample : P5299-01
 Misc : 6.95g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SB-01

Quant Time: Dec 19 02:51:25 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y121724S.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 18 05:27:13 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.707	168	233930	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.609	114	408249	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.414	117	351864	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.346	152	141411	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.061	65	125606	49.026	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	98.060%
35) Dibromofluoromethane	7.634	113	131817	45.814	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	91.620%
50) Toluene-d8	10.103	98	486937	48.352	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	96.700%
62) 4-Bromofluorobenzene	12.401	95	152376	38.680	ug/l	0.00
Spiked Amount	50.000	Range	29 - 146	Recovery	=	77.360%

Target Compounds Qvalue

 (#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY121824\
Data File : VY020642.D
Acq On : 18 Dec 2024 17:38
Operator : SY/MD
Sample : P5299-01
Misc : 6.95g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SB-01

Quant Time: Dec 19 02:51:25 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y121724S.M
Quant Title : SW846 8260
QLast Update : Wed Dec 18 05:27:13 2024
Response via : Initial Calibration

