



Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY122022\
 Data File : VY011910.D
 Acq On : 20 Dec 2022 09:20
 Operator : KP/MD
 Sample : BFB
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y122022S.M
 Title : SW846 8260
 Last Update : Tue Dec 20 12:37:37 2022



AutoFind: Scans 1748, 1749, 1750; Background Corrected with Scan 1740

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	17.6	12682	PASS
75	95	30	60	51.6	37133	PASS
95	95	100	100	100.0	71928	PASS
96	95	5	9	7.1	5130	PASS
173	174	0.00	2	1.0	616	PASS
174	95	50	100	88.4	63565	PASS
175	174	5	9	7.2	4592	PASS
176	174	95	101	95.8	60917	PASS
177	176	5	9	6.7	4110	PASS