

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY122023\
 Data File : VY016737.D
 Acq On : 21 Dec 2023 03:47
 Operator : SY/MD
 Sample : VY1220SBL01
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VY1220SBL01

Quant Time: Dec 21 04:10:31 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y122023S.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 21 01:49:45 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.813	168	132440	50.000	ug/l	0.01
34) 1,4-Difluorobenzene	8.716	114	247094	50.000	ug/l	0.01
63) Chlorobenzene-d5	11.514	117	212574	50.000	ug/l	0.01
72) 1,4-Dichlorobenzene-d4	13.446	152	80299	50.000	ug/l	0.01

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.167	65	81889	67.731	ug/l	0.01
Spiked Amount	50.000	Range	50 - 163	Recovery	=	135.460%
35) Dibromofluoromethane	7.746	113	83363	58.896	ug/l	0.02
Spiked Amount	50.000	Range	54 - 147	Recovery	=	117.800%
50) Toluene-d8	10.203	98	282641	55.394	ug/l	0.01
Spiked Amount	50.000	Range	58 - 134	Recovery	=	110.780%
62) 4-Bromofluorobenzene	12.501	95	86409	52.199	ug/l	0.01
Spiked Amount	50.000	Range	30 - 143	Recovery	=	104.400%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY122023\
Data File : VY016737.D
Acq On : 21 Dec 2023 03:47
Operator : SY/MD
Sample : VY1220SBL01
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VY1220SBL01

Quant Time: Dec 21 04:10:31 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y122023S.M
Quant Title : SW846 8260
QLast Update : Thu Dec 21 01:49:45 2023
Response via : Initial Calibration

