

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY122023\
 Data File : VY016744.D
 Acq On : 21 Dec 2023 06:44
 Operator : SY/MD
 Sample : 05876-13
 Misc : 4.36g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 COLX8

Quant Time: Dec 21 07:12:50 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y122023S.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 21 01:49:45 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.819	168	101207	50.000	ug/l	0.02
34) 1,4-Difluorobenzene	8.716	114	295922	50.000	ug/l	0.01
63) Chlorobenzene-d5	11.514	117	328260	50.000	ug/l	0.01
72) 1,4-Dichlorobenzene-d4	13.447	152	125850	50.000	ug/l	0.01
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.173	65	95283	103.130	ug/l	0.02
Spiked Amount	50.000	Range	50 - 163	Recovery	= 206.260%#	
35) Dibromofluoromethane	7.752	113	107577	63.463	ug/l	0.02
Spiked Amount	50.000	Range	54 - 147	Recovery	= 126.920%	
50) Toluene-d8	10.203	98	390211	63.857	ug/l	0.01
Spiked Amount	50.000	Range	58 - 134	Recovery	= 127.720%	
62) 4-Bromofluorobenzene	12.502	95	135596	68.397	ug/l	0.01
Spiked Amount	50.000	Range	30 - 143	Recovery	= 136.800%	
Target Compounds						
20) Methylene Chloride	4.765	84	10942	5.824	ug/l	Qvalue 96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY122023\
Data File : VY016744.D
Acq On : 21 Dec 2023 06:44
Operator : SY/MD
Sample : 05876-13
Misc : 4.36g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 29 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
COLX8

Quant Time: Dec 21 07:12:50 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y122023S.M
Quant Title : SW846 8260
QLast Update : Thu Dec 21 01:49:45 2023
Response via : Initial Calibration

